

BLAINE AIR PERMEABILITY (FINENESS) APPARATUS

Model Number: GOTT-BAPFA -01



DESCRIPTION

GOTT-BAPFA -01 is designed to determine the fineness of Portland cement in terms of the specific surface expressed as total surface area in square centimeters per gram of cement. The apparatus is supplied with glass U-tube manometer with valve, steel stand, test cell with disk and plunger all in stainless steel, rubber aspirator bulb, 1000 filter paper disks, manometric liquid, vaseline grease for better coupling tube/cell, funnel, brush.

ACCESSORIES

CODE STANDARD REFERENCE CEMENT 300-350

Standard reference cement 114q. to calibrate the blaine.

CODE **GLASS THERMOMETER** 300-351

-10 to +50 °C.

Manuals:

- (1) All manuals are written in English
- (2) Model Answer
- (3) Teaching Manuals

General Terms:

- (1) Accessories will be provided where applicable.
- (2) Manuals & Training will be provided where applicable.
- (3) Designs & Specifications are subject to change without notice.
- (4) We reserve the right to discontinue the manufacturing of any product.

Warranty:

2 Years

ORDERING INFORMATION:

ITEM	MODEL NUMBER	CODE
BLAINE AIR PERMEABILITY (FINENESS) APPARATUS	GOTT-BAPFA -01	300-349